

Quad Helix



Description/indication: The Quad Helix is a prefabricated expansion arch made of CHROMIUM wire ($\varnothing$ 0,9 mm) for attachment to the upper molars in palatal sheaths.

Contraindication: Product contains nickel and chromium. In case of allergies to these components, the Quad Helix should not be used.

Application: Make sure to ligate the Quad Helix firmly into the lingual/palatal sheaths using steel or elastic ligatures.

Note: Only to be used for the above indication and by qualified personnel.
For single use. Serious incidents in connection in connection with the use of the product must be reported immediately to the manufacturer and the authority.

DE/GB 1.000/09/26 G REF PM0115.01



CE 0044



SCHEU-DENTAL GmbH
www.scheu-dental.com

Am Burgberg 20
58642 Iserlohn · Germany

phone +49 2374 9288-0
fax +49 2374 9288-90

SCHEU